

Description: Microtox[®] Acute Reagent

Part Number: AZF686018A

Lot Number: 12A4000

Expiration Date: January 2014

Storage Condition: Frozen, -20 °C

Quality Control Testing

Test	Results	Specification	Pass/Fail
Phenol	EC50 (5 min) = 17.3 mg/L	13.0-26.0 mg/L	Pass
Zinc Sulfate	EC50 (15 min) = 4.5 mg/L	3.0-10.0 mg/L	Pass
Zinc ++ (ion) *Calculated	EC50 (15 min) = 1.0 mg/L	0.6-2.2 mg/L	Pass
DeltaTox Adjusted Light Reading	114	≥ 30	Pass

The values above are the mean of all Quality Control results obtained for this lot.
 All results are obtained using the Quality Control Procedure with an approved lot of
 Microtox 2% Diluent and Reconstitution Solution on the Model 500 and DeltaTox
 Analyzers.

Reviewed by: 
22 Feb 12